


200-Turn Field Coil

EM-6711



Included Equipment	Part Number
200-Turn Field Coil	EM-6711
Related Equipment	
500-Turn Field Coil	EM-6723
Detector Coil (400 turn)	EM-6712
Detector Coil (2000 turn)	EM-6713
Helmholtz Base	EM-6715
Recommended Equipment to Drive Coil (one needed)	
<i>ScienceWorkshop</i> 750 Interface	CI-7650
GLX Power Amplifier	PS-2006
Low Voltage AC/DC Power Supply	SF-9584A
Digital Function Generator/Amplifier	PI-9587C

Introduction

PASCO model EM-6711 is a 200-turn wire coil on a ring-shaped bobbin with a diameter of about 21 cm. The base of the assembly has three binding posts for connecting 4 mm banana plugs. To connect a device directly to the coil, use the two white binding posts. Do not allow the current to exceed 2 A.

A 1200 Ω resistor (located inside the base) is connected in series with the coil. Use the black binding post to include this resistor in a circuit.

A 1/4-20 threaded mounting hole in the base allows the coil to be fastened to other equipment.

Holes in the sidewalls of the bobbin allow students to see how the coil is wound. The holes are tangent with the inside of the winding so that students can measure the inner diameter.

Experiments

In combination with other equipment, the 200-Turn Field Coil can be used in a variety of demonstrations and experiments, including the following.

- Drive the field coil with a triangle wave at about 500 Hz. Use an oscilloscope to measure the emf induced in a detector coil. Move the detector coil to find the strength and direction of the magnetic field at various locations. Measure how the induced emf depends on the angle between the field coil and the detector coil. Vary the driving frequency to measure how the induced emf depends on the time rate of change of the magnetic field.

For details of this experiment see Christopher C. Jones, Faraday's Law apparatus for the freshman laboratory, *American Journal of Physics* 1987; 55 (12): 1148–1150.

- Do the above experiment with a *ScienceWorkshop* 750 interface and a voltage sensor. Connect the field coil directly to the interface output. (Do not use the 1200 Ω resistor; the driving current must be high, and the interface can measure

its own output current.) Use the voltage sensor to measure the induced emf in the detector coil.

- Drive the coil with direct current. Use a magnetic field sensor to find the strength and direction of the field at various locations.
- Mount two field coils on the Helmholtz Base (EM-6715) to create a Helmholtz coil.

Specifications

Turns	200
Wire	copper, 22 AWG (0.64 mm diameter)
Inner radius	10.06 cm
Outer radius	10.58 cm
Coil width	1.6 cm
Body material	Polycarbonate plastic

Technical Support

For assistance with any PASCO product, contact PASCO at:

Address: PASCO scientific
10101 Foothills Blvd.
Roseville, CA 95747-7100

Phone: 916-786-3800 (worldwide)
800-772-8700 (U.S.)

Fax: (916) 786-7565

Web: www.pasco.com

Email: support@pasco.com

Limited Warranty For a description of the product warranty, see the PASCO catalog.

Copyright The PASCO scientific 012-06176D *200-Turn Field Coil Instruction Sheet* is copyrighted with all rights reserved. Permission is granted to non-profit educational institutions for reproduction of any part of this manual, providing the reproductions are used only in their laboratories and classrooms, and are not sold for profit. Reproduction under any other circumstances, without the written consent of PASCO scientific, is prohibited.

Trademarks PASCO, and PASCO scientific are trademarks or registered trademarks of PASCO scientific, in the United States and/or in other countries. All other brands, products, or service names are or may be trademarks or service marks of, and are used to identify, products or services of, their respective owners. For more information visit www.pasco.com/legal.